

Tool 17: Disinfectant Selection Guide

Disinfection Guidelines are posted on the PICNET Website at:

<https://www.picnet.ca/guidelines/residential-care/>

NOTES:

- Be sure that the disinfectant product has a DIN number
- Check manufacturers information to ensure that product is effective against organisms in question
- Follow product instructions for dilution and contact time
- Unless otherwise stated on the product, use a detergent to clean surface of all visible debris prior to application of the disinfectant

Refer to the following [website](http://www.fraserhealth.ca/equipmentcleaning) to access the full version of the below Equipment Cleaning poster:

www.fraserhealth.ca/equipmentcleaning

The poster outlines cleaning and disinfection requirements of shared resident care equipment before and after use.

[Equipment Cleaning Poster](#)



The poster features four blue icons: a person using a walker, a vital signs monitor, a person in a wheelchair, and a person in a wheelchair being pushed. Below the icons is the title 'Clean shared resident care equipment before and after use' and a list of instructions. The Fraser Health logo is at the bottom.

Clean shared resident care equipment before and after use

- Disinfect **shared** resident care equipment **before** and **after** each resident use, including:
 - Walkers
 - Wheelchairs
 - Lifts
 - Vital signs machines
 - Thermometers
 - Stethoscopes
- Use hospital-grade disinfectant wipes

© Fraser Health 2021 